Appl. S.N. 09/839,940 Amdt. Dated August 27, 2003 Reply to Office Action of May 27, 2003

The listing of claims will replace all prior versions, and listings, of claims in the application:

Listing of Claims:

1. (previously amended) A radio frequency (RF) coil assembly for a very high field Magnetic Resonance Imaging (MRI) system comprising:

a plurality of conductors arranged cylindrically and disposed about a patient bore tube of the MRI system, said conductors of a selected length and having a width selected for said RF coil assembly to resonate at substantially high frequencies and to minimize conductor inductance; and,

a plurality of capacitive elements for electrically interconnecting said plurality of conductors at respective ends of said conductors.

2. (original) The RF coil assembly of claim 1 wherein the width of the conductors is selected in accordance with:

$$w_{\text{max}} = 2\pi * A/N$$

where w_{max} is the maximum width, A is the outer diameter radius of said patient bore tube and N is the number of said conductors.

- 3. (original) The RF coil assembly of claim 1 wherein said substantially high frequencies occurs in a range between about 64MHz to about 500 MHz.
- 4. (original) The RF coil assembly of claim 2 wherein said width is about 7.9 cm, and said number of conductors is 16.
- 5. (original) The RF coil assembly of claim 1 wherein said very high field MRI system produces a magnetic field of about 3 Tesla (3 T).
- 6. (original) The RF coil assembly of claim 1 wherein said plurality of conductors have a selectable length.
- 7. (original) The RF coil assembly of claim 6 wherein said selectable length is about 55 cm.
- 8. (original) The RF coil assembly of claim 1 wherein said capacitive elements are low inductance end ring capacitors.

RD-29,211

Appl. S.N. 09/839,940 Amdt. Dated August 27, 2003 Reply to Office Action of May 27, 2003

- 9. (original) The RF coil assembly of claim 1 wherein said conductors further include segmented slots for reducing eddy currents induced by gradient coils of said MRI system.
- 10. (original) The RF coil assembly of claim 1 further comprising a plurality of gaps disposed between said conductors.
- 11. (previously amended) A very high field Magnetic Resonance Imaging (MRI) system comprising:

a radio frequency (RF) coil assembly adapted to resonate at substantially high frequencies, the RF coil assembly having a plurality of conductors of selected length and selected width to minimize inductance;

- a RF coil shield assembly adapted to further reduce the inductance of the conductors contained within the RF coil assembly; and,
- a RF drive cable assembly adapted to electrically connect to the RF coil assembly.
- 12. (original) The MRI system of claim 11 wherein said substantially high frequencies occur in a range of about 64 MHz and about 500 MHz.
- 13. (original) The MRI system of claim 11 wherein said very high field MRI system produces a magnetic field of about 3 Tesla (3 T).
- 14. (original) The MRI system of claim 11 wherein said RF coil assembly comprises:

a plurality of conductors arranged cylindrically and disposed about a patient bore tube of the MRI system, said conductors having a width selected for said RF coil assembly to resonate at substantially high frequencies; and,

a plurality of capacitive elements for electrically interconnecting said plurality of conductors at respective ends of said conductors.

- 15. (original) The MRI system of claim 11 wherein said plurality of conductors and plurality of capacitive elements are adapted to form a band pass RF coil assembly configuration.
- 16. (original) The MRI system of claim 11 wherein said plurality of conductors and plurality of capacitive elements are adapted to form a low pass RF coil assembly configuration.
- 17. (original) The MRI system of claim 11 wherein said plurality of conductors and plurality of capacitive elements form a high pass RF coil assembly configuration.